

Product: VistaLab Pipette Tips
Packaging: Sample
Catalog Number: Sample

Lot Number: Sample
Manufacture Date: MM/DD/YYYY
Date of Expiry: MM/DD/YYYY

Tip material: Polypropylene

Filter material: Polyethylene

Sterilization: 10^{-6} SAL

Pre-sterilization is by electron beam radiation with post-sterilization verification. This lot has been processed and tested in accordance with ISO 11137-2. Product samples are bio-burdened.

RNase/DNase: RNase 10^{-9} Kunitz units/ μ L, DNase 10^{-7} Kunitz units/ μ L

This product has been tested and is free of any detectable RNase and DNase.

Endotoxin: 0.06 EU/mL

This product has been tested and meets the USP <85> and ANSI/AAMI ST72:2011 guidelines and FDA regulations of endotoxin limit for medical devices at 0.06 EU/mL.

Animal Content:

Resins, inks and packaging materials are not of animal origin and do not contain, make use of, or involve, at any point in the manufacturing process, raw materials of animal origin, including animal proteins.

VistaLab Technologies, Inc. is registered ISO 9001:2015

Signed on behalf of the Quality Department:



Jeff Calhoun
V.P. Product Development