IMPORTANT! Please Read



Note-

1. The motor will activate at seemingly random times when pressing the aspirate or dispense buttons. Don't worry, It's normal!

2. ali-O 2 will go into sleep mode after 3 minutes of non-use but will wake with any movement of the unit.

For more information on ali-Q's intelligent measuring system, see Operating Manual at www.vistalab.com/support

Aliquoting/repeat pipetting



ali-Q 2 VS model only:

Set desired aliquoting speed using speed dial

aliquot dials

ali-Q 2 LS

ali-Q 2 & ali-Q 2 VS

aliquot button



All ali-O 2 models:

- 1. Set aliquot volume on aliquot dial.
- 2. Aspirate adequate volume for total number of desired aliquots.
- 3. Press and HOLD the purple 🚠 aliquot button until the aliquot is dispensed. You will hear and feel the motor come on. Repeat press and hold for each aliquot. For maximum precision, discard first aliquot.

Note: if you don't hold the button long enough for the full aliquot volume, you will hear an "alert" sound and red light will flash.



Description

ali-Q 2 pipet controller ali-Q 2 LS pipet controller ali-Q 2 VS pipet controller VistaLab Catalog #

2100-0005 2100-0003 2100-1005

Additional box contents:

spare 0.22µm membrane filter power supply w/outlet adapters wall mount calibration certificate 9020-4003 9020-9005 9020-9008

Register your ali-Q 2 online at www.vistalab.com

for 1 year Warranty

Go to www.vistalab.com/support for full Operating Manual



Complete your optimal liquid handling system with Wobble-not™ serological pipets VistaLab Catalog

Standard Length

5 mL	(200/cs)
10 mL	(200/cs)
25 mL	(200/cs)
50 mL	(100/cs)

Shorty Length

10 mL (200/cs) 25 mL (100/cs)



4090-0005 4090-0010 4090-0025 4090-0050

5090-0010 5090-0025

2 Geneva Road Brewster, NY 10509 USA (888) 652-6520; (914) 244-6226 customer.service@vistalab.com ©2020 VistaLab Technologies, Inc. All rights reserved.