

PIPETTE IDENTIFICATION

CUSTOMER ID	JVW 25-850	CATALOG NO.	1060-0850
CUSTOMER NAME	[REDACTED]	MODEL	Ovation 12-Channel Multic
ADDRESS 1	[REDACTED]	SERIAL NO.	213712
ADDRESS 2	[REDACTED]	REPAIR NO.	WD37542
CITY, STATE, ZIP	Hayward CA 94545	TECHNICIAN ID	Y.B.

TESTING CONDITIONS

WATER TEMP (°C)	21.02	BALANCE MODEL	Sartorius MC210S
AIR TEMP (°C)	21.70	BALANCE SERIAL NO.	70608696
RELATIVE HUMIDITY (%)	52.3	TIPS USED	VistaLab 4060-3004
BAROMETRIC PRESSURE (mmHg)	746.70	"Z" FACTOR	1.0031
		TEST METHOD	ISO 8655

TESTING INFORMATION

TESTING DATA

(measured weight - mg)	Volume 1	Volume 2
Weighing 1	85.55	849.54
Weighing 2	85.51	849.47
Weighing 3	85.45	849.45
Weighing 4	85.27	849.64
Weighing 5	85.40	849.81
Weighing 6	85.64	849.69
Weighing 7	85.52	849.69
Weighing 8	85.72	849.38
Weighing 9	85.48	849.56
Weighing 10	85.46	849.78
Mean (mg)	85.50	849.60

CALIBRATION VERIFICATION RESULTS

Mean Test Volume (µL)	85.77	852.23
Accuracy (±%)	0.91	0.26
Precision (CV%)	0.14	0.02
Uncertainty (± µL)	0.23	0.87

PERFORMANCE SPECIFICATIONS

Nominal Test Volume (µL)	85.00	850.00
Accuracy (±%)	3.50	1.00
Precision (CV%)	1.00	0.30

PIPETTE STATUS PASS PASS

Comments:

CALIBRATION RESULTS FOR CHANNEL #11.

Performed By: Yeha Balal Title: Technician Date of Calibration: 4/11/2025